

Smokerlyzer® CO Chart



COppm

%COHb

30	5.43
29	5.27
28	5.11
27	4.95
26	4.79
25	4.63
24	4.47
23	4.31
22	4.15
21	3.99
20	3.83
19	3.67
18	3.51
17	3.35
16	3.19
15	3.03
14	2.87
13	2.71
12	2.55
11	2.39
10	2.23
09	2.07
08	1.91
07	1.75
06	1.59
05	1.43
04	1.27
03	1.11
02	0.95
01	0.79

Reference:

1. COppm- %COHb calculation taken from: Jarvis M et al (1986) "low cost Carbon Monoxide monitors in smoking assessment." Thorax 41 pp 886-887. Issue 4- November 2015, Part No: LAB469

www.covita.net | (800) 707-5751